

تست	قابل قبول DEV%
Glucose	15.4
Creatinine	17.8
Urea	11.4
Uric Acid	15.4
Cholesterol	15.2
Triglyceride	15.2
HDL	24
LDL	30
Calcium	8
Phosphate	15.6
Sodium	3.2
Potassium	5.8
Lithium	22
ALP	31
AST	25
ALT	34
CPK	37
LDH	16
Direct Bilirubin	38.4
Total Bilirubin	38.4
Iron	30
TIBC	30
UIBC	30
Albumin	15
Protein	7.8
Lipase	36
Amylase	23
GGT	31.4
Magnesium	20
Transferrin	30
Ferritin	20
T4	30
T3 Uptake	12
T3	30
TSH	30
Total PSA	20
hCG	20
HbA1c	20
Urine Protein	20
Troponin T	22
Troponin I	22
CK-MB activity	20
CK-MB mass	20

تست	قابل قبول DEV%
WBC	16
Rbc	6
HB	6
HCT	8
MCV	8
MCH	6
MCHC	8
PLT	20
PT	18
PTT	18

Deviation



بیمارستان گلگاه  
شرکت بین کیفیت آزمایشگاه های پزشکی